

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0260856	(X3) Date Survey Completed 02/16/2023
Name of Provider or Supplier Associated Pathologists, Llc	Street Address, City, State 1901 Phoenix Blvd, Suite 210, Atlanta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on February 16, 2023. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Standard Operating Procedure (SOP) and staff interview, the laboratory failed to establish a safety procedure for the Eyewash. The Findings include: 1. SOP document review revealed that the laboratory failed to establish a safety procedure for the sink eyewash in the laboratory. 2. No maintenance was being performed on the Eyewash for 2021, 2022, and thus far 2023 (January-December). 3. During an interview with the Director of Operations on February 16, 2023, at 12:50 PM, in the conference room, confirmed that the laboratory failed to establish a safety procedure for the Eyewash and no maintenance was performed on the Eyewash for 2021, 2022, and thus far 2023.</p>