

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0261382	(X3) Date Survey Completed 12/13/2023
Name of Provider or Supplier Longstreet Clinic	Street Address, City, State 725 Jesse Jewel Parkway, Gainesville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	On January 03, 2024 an off site follow-up review was completed. The report revealed that the plan of correction was found to be acceptable. The facility is now in compliance with CLIA regulations.
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on staff interview and lab tour, the laboratory had only one running water "dirty" sink and failed to install an eyewash station in the laboratory to protect against accidental chemical and biohazard facial contamination, as required by CLIA for years 2022 - 2023, in building #715, suite 390. Findings: 1. Review of maintenance records revealed that the laboratory failed to have eyewash inspection documents and failed to have eyewash flush equipment installed in the laboratory to protect against accidental biological or chemical facial contamination in the lab from 2022 - 2023, in building #715, suite 390. 2. Interviews with staff and Technical Consultant (TC) (TP #3 CMS 209) at 2:00 PM on 12/13/2023 confirmed the laboratory failed to have eyewash inspection records and did not have an eyewash installed, at the time of survey, 12/13/2023.</p>