

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0261672	(X3) Date Survey Completed 10/09/2018
Name of Provider or Supplier Chatuge Regional Hosp & Nursing Home	Street Address, City, State 110 Main Street, Hiawassee, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review was completed on October 9, 2018. At the time of the review, the laboratory was not in compliance with the Clinical Laboratory Improvement Amendments of 1988, 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on proficiency testing desk review using the Centers for Medicare and Medicaid (CMS) Casper Reports 155 and 153 and review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) report, the laboratory failed to maintain satisfactory performance in event 1 of 2018 resulting in the first

	unsuccessful occurrence for ABO blood group analyte # 865, Rhesus type D(RHO) analyte # 875, and compatibility testing, analyte # 895. Findings include: Refer to D 2155
D2155	ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(c) Failure to participate in a testing event is unsatisfactory performance and results in a score of 0 for the testing event. Consideration may be given to those laboratories failing to participate in a testing event only if-- (1) Patient testing was suspended during the time frame allotted for testing and reporting proficiency testing results; (2) The laboratory notifies the inspecting agency and the proficiency testing program within the time frame for submitting proficiency testing results of the suspension of patient testing and the circumstances associated with failure to perform tests on proficiency testing samples; and (3) The laboratory participated in the previous two proficiency testing events. This STANDARD is not met as evidenced by: Based on proficiency testing desk review using the Centers for Medicare and Medicaid (CMS) Casper Reports 155 and 153 and review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) report, the laboratory failed to maintain satisfactory performance in event 2 of 2018 resulting in the first unsuccessful occurrence for ABO blood group analyte # 865, Rhesus type D(RHO) analyte # 875, and compatibility testing, analyte # 895. Findings include: 1. Desk review of Casper Reports 153 and 155 disclosed the laboratory failed analytes #865 ABO, #875 D(RHO) and #895 compatibility testing on event 1 of 2018 with a score of 0%. 2. Desk review of the laboratory's proficiency testing reports from API confirmed the laboratory failed the three analytes listed above on event 1 of 2018 for failure to participate, resulting in unsuccessful performance.
D6000	MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403 The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart. This CONDITION is not met as evidenced by: Based on an in-office desk review of proficiency testing records, the laboratory director failed to ensure the laboratory maintained compliance with successful performance in event 1 of 2018 resulting in unsuccessful occurrence for ABO blood group analyte # 865, Rhesus type D(RHO) analyte # 875, and compatibility testing, analyte # 895. Findings include: Refer to D 6016
D6016	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory

director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;

This STANDARD is not met as evidenced by:

Based on proficiency testing desk review using the Centers for Medicare and Medicaid (CMS) Casper Reports 155 and 153 and review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) report, the laboratory director failed to ensure the laboratory maintained satisfactory performance in event 1 of 2018 resulting in unsuccessful occurrence for ABO blood group analyte # 865, Rhesus type D(RHO) analyte # 875, and compatibility testing, analyte # 895. Findings include: 1. Desk review of Casper Reports 153 and 155 disclosed the laboratory failed analytes #865 ABO, #875 D(RHO) and #895 compatibility testing on event 1 of 2018 with a score of 0%. 2. Desk review of the laboratory's proficiency testing reports from API confirmed the laboratory failed the three analytes listed above on event 1 of 2018 for failure to participate, resulting in unsuccessful performance.