

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0264038	(X3) Date Survey Completed 08/19/2026
Name of Provider or Supplier Primary Pediatrics Of Macon	Street Address, City, State 5300 Bowman Road, Macon, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on August 19, 2026. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D3011	FACILITIES CFR(s): 493.1101(d) Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials. This STANDARD is not met as evidenced by: Based on observation during the initial tour of the laboratory, a large fruit bowl was observed in the refrigerator where vaccines, hematology reagents and controls were stored on the day of survey August 19, 2026. Findings: 1.) During the tour of the laboratory, a large fruit bowl from Chik-Fil-A was in the vaccines and reagent refrigerator on Smith's hall lab. It was brought to the attention of the office manager. There was no laboratory staff that claimed ownership of the food item in that refrigerator on the day of survey August 19, 2026. 2.) Interviews with laboratory staff and office manager on Smith's hall on August 19, 2026 at approximately 10:30 AM confirmed the presence of the food item in the laboratory refrigerator.