Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 11D0264744	**(X3) Date Survey Completed** 11/03/2020
Name of Provider or Supplier Savannah Primary Care Associates, Llc	**Street Address, City, State** 4451 Paulsen Street, Savannah, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on November 3, 2020. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5429	MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1) For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer. This STANDARD is not met as evidenced by: Based on equipment maintenance document review, observation, and staff interview, the laboratory failed to perform and document equipment maintenance as required. Findings include: 1. Equipment maintenance document review revealed the lack of daily maintenance logs for the Accuscope microscope for the following dates: 2018 -- October through December, 2019, and 2020 thus far. 2. Observation during the laboratory tour on 11/3/2020 at approximately 1:30 p.m. revealed the following: Hamilton Bell centrifuge last calibrated 6/2018; Fisher-Scientific vortex mixer was due for calibration 6/2019; Fisher-Scientific AccuSpin 24C centrifuge last calibrated 6 /2018. 3. An interview with the laboratory supervisor in the laboratory on 11/3/2020 at approximately 1:30 p.m. confirmed the lack of microscope daily maintenance logs for the aforementioned dates. During the same interview, the laboratory supervisor confirmed the lack of required calibration for the aforementioned centrifuges and vortex mixer.