

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0663100	(X3) Date Survey Completed 11/02/2020
Name of Provider or Supplier Vidalia Medical Associates	Street Address, City, State 200 Maple Drive, Vidalia, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on November 20, 2020. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on laboratory maintenance document review and staff interview, the laboratory failed to observe safety procedures to ensure protection from physical chemical and biohazardous materials as required. Findings include: 1. Laboratory maintenance document review revealed the lack of eyewash equipment maintenance documents available at the time of survey for the following dates: 2018 -- September through December; 2019, and 2020 thus far. 2. An interview with the laboratory manager in the conference room on 11/2/2020 at approximately 4:45 p.m. confirmed the lack of eyewash equipment maintenance for the aforementioned dates.</p>