

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0675407	(X3) Date Survey Completed 08/03/2022
Name of Provider or Supplier Georgia Dermatology Partners	Street Address, City, State 2383 Pate Street, Snellville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on August 03, 2022. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5781	<p>CORRECTIVE ACTIONS CFR(s): 493.1282(b)(1)</p> <p>(b) The laboratory must document all corrective actions taken, including actions taken when any of the following occur: (b)(1) Test systems do not meet the laboratory's verified or established performance specifications, as determined in 493.1253(b), which include but are not limited to-- (b)(1)(i) Equipment or methodologies that perform outside of established operating parameters or performance specifications; (b)(1)(ii) Patient test values that are outside of the laboratory's reportable range of test results for the test system; and (b)(1)(iii) When the laboratory determines that the reference intervals (normal values) for a test procedure are inappropriate for the laboratory's patient population.</p> <p>This STANDARD is not met as evidenced by: Based on maintenance records review and interview with the laboratory director, the laboratory failed to document corrective action for out of range values of Relative Humidity (RH) in 2020 and 2022. Findings include: 1.) Maintenance records review of revealed Maximum Relative Humidity(RH) on logs is 60%. The following days; 06/23/2020(66%), 06/29/2022(61%), 07/27/2022(61%), 08/10/2022, 08/28/2022 and 08/31/2022(61%), 09/07/2022 to 09/09/2022(65%,65%, 62%), 09/15/2022 to 09/16/2022 (61%,64%,65%) and 07/09/2022(67%) 2.) An interview with the laboratory director TP#1 (CMS 209) at approximately 11:45 am in his office on 08/03/2022 confirmed no corrective action was documented for the out of range Humidity values in 2020 and 2022.</p>