Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0682811	(X3) Date Survey Completed 03/28/2018
Name of Provider or Supplier William E Freeman Md	**Street Address, City, State** 136 S Houston Lake Dr, Warner Robins, GA	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on March 28, 2018. The laboratory was not in compliance with all applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D2000	ENROLLMENT AND TESTING OF SAMPLES CFR(s): 493.801 Each laboratory must enroll in a proficiency testing (PT) program that meets the criteria in subpart I of this part and is approved by HHS. The laboratory must enroll in an approved program or programs for each of the specialties and subspecialties for which it seeks certification. The laboratory must test the samples in the same manner as patients' specimens. For laboratories subject to 42 CFR part 493 published on March 14, 1990 (55 FR 9538) prior to September 1, 1992, the rules of this subpart are effective on September 1, 1992. For all other laboratories, the rules of this subpart are effective January 1, 1994. This CONDITION is not met as evidenced by: Based on Proficiency Testing documentation for the year 2016, 2017 and 2018, and staff interview, the laboratory failed to enroll in a Proficienccy Testing(PT) program for the Potassium Hydrozide Preparation(KOH) and Dematophyte Test Media (DTM) being performed.. Reference: 2001
D2001	ENROLLMENT CFR(s): 493.801(a)(1)(2)(i) The laboratory must-- (1) Notify HHS of the approved program or programs in which it chooses to participate to meet proficiency testing requirements of this subpart. (2)(i) Designate the program(s) to be used for each specialty, subspecialty, and analyte or test to determine compliance with this subpart if the laboratory participates in more

	than one proficiency testing program approved by CMS; This STANDARD is not met as evidenced by: Based on Proficiency Testing documentation for 2016, 2017, and 2018, and staff interview the laboratory failed to enroll in a Proficiency testing program for Potassium Hydroxide Preparation (KOH) and Dermatophyte Test Medium (DTM). Findings: 1. Staff interview with staff # 1 (CMS 209 form) on March 28, 2018, in the Surgical Center Nursing Station, confirmed that the Laboratory did not enroll in a Proficiency program for KOH and DTM testing.