

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0684378	(X3) Date Survey Completed 02/09/2021
Name of Provider or Supplier Wellstar Medical Group West Cobb Med Ctr	Street Address, City, State 3707 Largent Way Nw, Marietta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on February 9, 2021. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5791	<p>ANALYTIC SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1289(a)(c)</p> <p>(a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and when indicated, correct problems identified in the analytic systems specified in 493.1251 through 493.1283. (c) The laboratory must document all analytic systems assessment activities.</p> <p>This STANDARD is not met as evidenced by: Based on review of quality assurance (QA) records and staff interview, the lab failed to document QA monitors per the written policies and procedures (SOP). Findings include: 1. Review of QA records of 2019 and 2020 revealed QA monitors were not performed in 2019; in January through September or November through December 2020 per the SOP. 2. Interview with staff #1 (CMS 209 form) on 02/09/2021 at approximately 2:30 PM in the conference room, confirmed the QA monitors were not documented for the aforementioned dates.</p>
D6021	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(5)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(5) Ensure that quality assessment programs are established and</p>

maintained to assure the quality of laboratory services provided.

This STANDARD is not met as evidenced by:

Based on review of the laboratory records and staff interview, the laboratory director failed to ensure that the quality assessment(QA) program is maintained to identify failures in quality. Findings include: 1. Review of the laboratory's QA records revealed the lack of documentation of QA monitors for 2019, January through September 2020 or November through December 2020. 2. Interview with staff #1 (CMS 209 form) on 02/09/2021 at approximately 2:30 PM in the conference room, confirmed the QA monitors were not documented for the aforementioned dates., confirmed the missing aforementioned QA.