

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 11D0726627 | (X3) Date Survey Completed 04/19/2018 |
| Name of Provider or Supplier Pediatric First, Pc | Street Address, City, State 1049 N Houston Rd, Warner Robins, GA | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D0000 | A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on April 19, 2018. The laboratory was not in compliance with all applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited: |
| D5439 | <p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on review of the calibration records for the Abbott Emerald Hematology</p> |

Analyzer, and staff interview the laboratory failed to calibrate the Abbott Emerald Hematology Analyzer at least every 6 months. Findings: 1. Review of the 2017 and 2018 calibration records for the Abbott Emerald Hematology Analyzer showed that the calibration was performed in January of 2017, and August of 2017 (7 months), and in March of 2018 (7 months). 2. Interview with the office manager on April 19, 2018 at 3:40pm in the Lab Director's office confirmed that the Abbott Emerald Hematology Analyzer was calibrated in January of 2017, and August of 2017 (7 months), and in March of 2018 (7 months).