

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D0924021	<b>(X3) Date Survey Completed</b>  06/04/2024
<b>Name of Provider or Supplier</b>  Morehouse School Of Medicine	<b>Street Address, City, State</b>  720 Westview Drive, Sw, Room F-126, Atlanta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on June 04, 2024. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
<b>D3011</b>	<p><b>FACILITIES</b> CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on document review and staff interview during the lab tour, the laboratory failed to document and review inspections of the flush eyewash equipment on a weekly basis in the laboratory to protect against accidental chemical and biohazard facial contamination as required by CLIA from July 2022 to June 2024. Findings: 1. A review of maintenance records revealed that the laboratory failed to document and review inspections of the eyewash flush equipment, on a weekly basis, to protect against accidental biohazard or chemical facial contaminations from July 2022 to June 2024. 2. Interviews with the General (GS) and Technical Supervisor (TS) staff (TP #2 and #3 CMS 209) at approximately 12:00 PM on 06/04/2024 confirmed the laboratory failed to document and review inspections of the flush eyewash equipment on a weekly basis from July 2024 to June 2024..</p>