

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0932951	(X3) Date Survey Completed 02/11/2022
Name of Provider or Supplier Youthcare Pediatrics Of Central Georgia Pc	Street Address, City, State 233 North Houston Road Suite 140 H, Warner Robins, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) Recertification survey was completed on February 11,2022. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by:</p>

Based on calibration document review and staff interview, the lab failed to calibrate the Cell-Dyn Emerald analyzer every six (6) months as required by the manufacturer. Findings include: 1. Review of calibration data revealed the Emerald was calibrated 1 /18/19. There was no documentation of a calibration after that date. The lab failed to calibrate the Emerald, in July/2019, January 2020, July 2020, January 2021, July 2021, and January 2022. 2. Interview with testing personnel #3, (CMS 209 form) and the Laboratory Director, on February 11, 2022 at approximately 12 PM in the Office Manager's office , confirmed the aforementioned statement.

D6004

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1407(a)(b)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (a) The laboratory director, if qualified, may perform the duties of the technical consultant, clinical consultant, and testing personnel, or delegate these responsibilities to personnel meeting the qualifications of 493.1409, 493.1415, and 493.1421, respectively. (b) If the laboratory director reapportions performance of his or her responsibilities, he or she remains responsible for ensuring that all duties are properly performed.

This STANDARD is not met as evidenced by:
Based on review of the Calibration Documents for the the Emerald Hematology Analyzer (Emerald), and staff interview, the Laboratory Director(LD) failed to provide oversight for the overall operation and administration of the laboratory. Findings: 1. Review of the Emerald Calibration documents, the lab did not perform a calibration on the Emerald every six months from 01/18/2019 to 02/11/2022 as required by the manufacturer. 2. Interview with staff #3 (CMS 209 form), and the LD, on 02/11/2022 at approximately 12:30 in the Office Manager's office confirmed that the Emerald was not calibrated after January 19, 2019.