

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0952100	(X3) Date Survey Completed 06/17/2019
Name of Provider or Supplier Family Physicians Of Evans	Street Address, City, State 465 North Belair Road, Suite 1c, Evans, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) Recertification survey was completed on June 17, 2019. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D2007	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods</p> <p>This STANDARD is not met as evidenced by: Based on proficiency test (PT) document review and staff interview, the laboratory failed to examine the PT samples with the laboratory's regular patient workload by personnel who routinely perform the moderate-complexity laboratory testing. Findings include: 1. American Proficiency Institute (API) document review revealed Staff #3 (CMS 209) performed all of the Hematology PT sample testing for Events 1, 2, and 3 of 2018. 2. An interview in a medical office on 6/17/19 with the technical consultant at approximately 3:45 p.m. confirmed the aforementioned PT events were performed by the same testing personnel.</p>