

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0981282	(X3) Date Survey Completed 03/21/2019
Name of Provider or Supplier Kids First Pediatrics	Street Address, City, State 530 North Cobb Street, Milledgeville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on March 21, 2019. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiency was cited:
D5221	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d)</p> <p>All proficiency testing evaluation and verification activities must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Academy of Family Physicians (AAFP) Proficiency Provider, and staff interview, the laboratory failed to document correction action taken for unsatisfactory scores. Findings: 1 Review of the AAFP evaluation report for 2018 event B, showed a score of 0% for the analyte Red Blood Cell,(RBC) and Hematocrit (HCT). There was no corrective action documented for the RBC and HCT score of 0%. 2. Interview with staff # 2(CMS 209 form), the Laboratory Manager, and the practice manager, on March 21, 2019 at approximately 11am in teh Practice Manager's office, confirmed that there was no documented corrective action for the RBC and HCT score of 0%.</p>