

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0982961	(X3) Date Survey Completed 09/26/2019
Name of Provider or Supplier East Georgia Diagnostic Services, Inc	Street Address, City, State 333 South Walnut Street, Statesboro, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on September 26, 2019. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5429	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1)</p> <p>For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on observation and staff interview, the laboratory failed to perform and document required equipment maintenance. Findings include: 1. Observation during the laboratory tour on September 26, 2019, at approximately 10:45 a.m. revealed there were no professional calibration records for the Nikon microscope for 2018 and 2019 thus far. 2. An interview with the lab director on September 26, 2019, in the office area at approximately 10:45 a.m. confirmed the lack of professional calibration documents for the Nikon microscope for the aforementioned time periods..</p>