

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0991500	(X3) Date Survey Completed 01/15/2025
Name of Provider or Supplier Dermatology Specialists Of West Ga	Street Address, City, State 109 Professional Place, Carrollton, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on 01/15/2025. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiency was cited:
D5791	<p>ANALYTIC SYSTEMS QUALITY ASSESSMENT CFR(s): 493.1289(a)(c)</p> <p>(a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and when indicated, correct problems identified in the analytic systems specified in 493.1251 through 493.1283. (c) The laboratory must document all analytic systems assessment activities.</p> <p>This STANDARD is not met as evidenced by: Based on lack of annual microscope maintenance document, review of the policy and procedure manual (SOP), and subsequent interview with the lab director, the laboratory failed to perform and document microscope annual maintenance. Findings include: 1. Review of the microscope maintenance documents of 2023, 2024, & 2025 revealed there was no annual maintenance document available for the Nikon ECLIPSE 50i or the Olympus CH-B145-T-2 microscopes for the year of 2024. 2. Review of the SOP revealed microscope maintenance was to be done weekly and annually. 3. An interview with the laboratory director on 1/15/25 in the breakroom at 12:00 p.m. confirmed the lack of annual maintenance for the laboratory microscopes for the year of 2024.</p>