

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D0997580	(X3) Date Survey Completed 04/12/2018
Name of Provider or Supplier Georgetown Pediatrics	Street Address, City, State 416 Pirkle Ferry Road Suite 300 J, Cumming, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An offsite revisit survey was conducted on May 16, 2018, for all previous deficiencies cited on 4/12/2018. All deficiencies have been corrected. The facility is in compliance with with all regulations surveyed.
D2007	<p>TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1)</p> <p>The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods</p> <p>This STANDARD is not met as evidenced by: Based on proficiency testing (PT) document review and staff interview, the laboratory failed to test PT samples with the laboratory's regular patient workload by testing personnel (TP) who routinely perform the testing in the laboratory. Findings include: 1. Medical Laboratory Evaluation (MLE) PT document review revealed Staff #4 (CMS 209) did not participate in testing of PT samples from 2016 through 2018 thus far. 2. An interview with Staff #2 (CMS 209) in a doctor's office on 4/12/18 at approximately 11:45 AM confirmed Staff #4 (CMS 209) did not participate in testing of PT samples from 2016 through 2018 thus far.</p>