

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D1001114	<b>(X3) Date Survey Completed</b>  01/17/2025
<b>Name of Provider or Supplier</b>  Gwinnett Clinic Ltd	<b>Street Address, City, State</b>  475 Philip Boulevard, Suite 301, Lawrenceville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on January 17, 2025. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
<b>D3011</b>	<p><b>FACILITIES</b> CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based upon a walkthrough of the facility and an interview with the TC, the eyewash maintenance and use standards were not met. FINDINGS: 1. Upon tour of the laboratory, it was confirmed that the eyewash was unable to maintain water flow without requiring the use of the operator's hands or remain operational without manually holding. 3. Observation of the eyewash station, during the lab tour, revealed that the eyewash was not large enough to provide room for the eyelids to be held open with the hands while the eyes are in the stream of water. 4. An interview, on January 17, 2025, at 1:00 pm., with the TC during the walkthrough and during the exit interview confirmed the findings above.</p>