

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D1016973	<b>(X3) Date Survey Completed</b>  11/25/2025
<b>Name of Provider or Supplier</b>  Aqua Dermatology Of Georgia , Pc	<b>Street Address, City, State</b>  835 Cogburn Avenue, Marietta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) Recertification survey was completed on November 25, 2025. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
<b>D6107</b>	<p><b>LABORATORY DIRECTOR RESPONSIBILITIES</b> CFR(s): 493.1445(e)(15)</p> <p>(e)(15) Specify, in writing, the responsibilities and duties of each consultant and each supervisor, as well as each person engaged in the performance of the preanalytic, analytic, and postanalytic phases of testing, that identifies which examinations and procedures each individual is authorized to perform, whether supervision is required for specimen processing, test performance or result reporting and whether supervisory or director review is required prior to reporting patient test results.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory policy and procedure manual (SOP) and staff interview, the laboratory director (LD) failed to specify, in writing the duties and responsibilities of each person engaged in the performance of the preanalytic, analytic, and postanalytic phases of laboratory testing. Findings include: 1. SOP review revealed the LD failed to specify in writing the duties and responsibilities of each person engaged in the performance of all phases of laboratory testing. No defined duties and responsibilities for the Lab Director, Clinical Consultant, Technical Supervisor, General Supervisor, or the Testing Personnel. 2. An interview with the clinical supervisor in the review room on 11/25/25 at 2:20 p.m. confirmed the SOP did not contain a duties and responsibilities policy and procedure for the aforementioned positions.</p>