

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D1024365	<b>(X3) Date Survey Completed</b>  11/16/2020
<b>Name of Provider or Supplier</b>  Kids Health	<b>Street Address, City, State</b>  1216 Sherwood Park Drive, Gainesville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	On December 22,2020, an off site followup review was completed. The report revealed that corrective action was found to be acceptable and corrected. The facility is now in compliance with with all regulations surveyed.
<b>D5807</b>	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by: Based on patient chart review and reports for 2019 and 2020, and laboratory director interview, the laboratory failed to provide the units of measure and reference ranges for results from the Coulter AcT Diff II hematology analyzer. Findings: 1. Review of patient charts and patient reports for 2019 and 2020, the reports did not have the units of measure and the reference range for each analyte of the Complete Blood Count, from the Coulter AcT Diff II Hematology Analyzer. 2. Interview with the Laboratory Director on 11/16/2020 at approximately 02:30 PM in the review room, confirmed that the units of measure and the reference ranges for each analyte were not on the reports pulled from 2019 and 2020, from the Coulter AcT Diff II Analyzer.</p>