

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D1032047	(X3) Date Survey Completed 01/26/2023
Name of Provider or Supplier Longstreet Clinic Pc, The	Street Address, City, State 4019 Executive Drive, Oakwood, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on January 26, 2023. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on observation and staff interview, the laboratory failed to ensure the blood agar plates were not used after their expiration date, 6 out of 6 blood agar plates were expired. The Findings include: 1. Observation during the laboratory tour on January 26, 2023 at approximately 11:00 AM revealed 6 blood agar plate were expired on January 17, 2023, and three expired blood agar plates had specimens plated on them. Observation during the same tour at 11:05 AM on January 26, 2023 revealed that three plates were discarded, but the plates that were plated with specimens on them remained, during the time of the survey. 2. During an interview with Technical Consultant(CMS-209) in the laboratory on January 26, 2023 at approximately 11:00AM, confirmed there were 6 total blood agar plates that were expired and 3 of the plates had specimens on them, during the laboratory tour.</p>