

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D1069247	<b>(X3) Date Survey Completed</b>  12/09/2025
<b>Name of Provider or Supplier</b>  Lane Dermatology And Dermatologic Surgery	<b>Street Address, City, State</b>  1210 Brookstone Centre Parkway, Columbus, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) Recertification Survey was completed on December 09, 2025. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
<b>D6168</b>	<p>TESTING PERSONNEL CFR(s): 493.1487</p> <p>The laboratory has a sufficient number of individuals who meet the qualification requirements of 493.1489 of this subpart to perform the functions specified in 493.1495 of this subpart for the volume and complexity of testing performed.</p> <p>This CONDITION is not met as evidenced by: Based on personnel record review from 2023 - 2025 and staff interview, the laboratory failed to ensure that all testing personnel met the minimum educational requirements for testing personnel (TP) in a High complexity laboratory, on the date of survey, December 09, 2025. Reference D6171.</p>
<b>D6171</b>	<p>TESTING PERSONNEL QUALIFICATIONS CFR(s): 493.1489(b)</p> <p>(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located; or (b)(2)(i) Have earned a doctoral, master's, or bachelor's degree in a chemical, biological, clinical or medical laboratory science, or medical technology from an accredited institution; or (b)(2)(ii) Be qualified under the requirements of 493.1443(b)(3) or 493.1449(c)(4) or (5); or (b)(3)(i) Have earned an associate degree in a laboratory science or medical laboratory technology from an accredited institution or (b)(3)(ii) Have education and training equivalent to that specified in paragraph (b)(2)(i) of this section that includes (b)(3)(ii)</p>

(A) At least 60 semester hours, or equivalent, from an accredited institution that, at a minimum, includes either (b)(3)(ii)(A)(1) 24 semester hours of medical laboratory technology courses; or (b)(3)(ii)(A)(2) 24 semester hours of science courses that include (b)(3)(ii)(A)(2)(i) 6 semester hours of chemistry; (b)(3)(ii)(A)(2)(ii) 6 semester hours of biology; and (b)(3)(ii)(A)(2)(iii) 12 semester hours of chemistry, biology, or medical laboratory technology in any combination; and (b)(3)(ii)(B) Have laboratory training that includes: (b)(3)(ii)(B)(1) Completion of a clinical laboratory training program approved or accredited by the ABHES or the CAAHEP (this training may be included in the 60 semester hours listed in paragraph (b)(3)(ii)(A) of this section); or (b)(3)(ii)(B)(2) At least 3 months documented laboratory training in each specialty in which the individual performs high complexity testing; or (b)(4) Successful completion of an official U.S. military medical laboratory procedures training course of at least 50 weeks duration and having held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); or (b)(5) Notwithstanding any other provision of this section, an individual is considered qualified as a high complexity testing personnel under this section if they were qualified and serving as a high complexity testing personnel in a CLIA-certified laboratory as of December 28, 2024, and have done so continuously since December 28, 2024. (b)(6) For blood gas analysis (b)(6)(i) Be qualified under paragraph (b)(1), (2), (3), (4), or (5) of this section; or (b)(6)(ii) Have earned a bachelor's degree in respiratory therapy or cardiovascular technology from an accredited institution; or (b)(6)(iii) Have earned an associate degree related to pulmonary function from an accredited institution. (b)(7) For histopathology, meet the qualifications of 493.1449 (b) or (f) to perform tissue examinations.

This STANDARD is not met as evidenced by:

Based on personnel record review and staff interviews, the laboratory failed to ensure testing personnel (TP) met the educational qualification requirements for a high complexity testing laboratory, on the date of survey, December 9, 2025. Findings: 1. A review of 2023 - 2025 personnel records revealed HT2, HT3, and HT4 ( See Form CMS-209) failed to meet the minimum education requirements for high complexity testing personnel, on the date of survey, 12/09/2025. 2. An exit interview with the laboratory director and staff, in the review room, on 12/09/2025, at approximately 12:00 PM , confirmed the laboratory failed to ensure all testing personnel met the minimum educational requirements for a high complexity laboratory as of 12/09/2025.