

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D1090583	<b>(X3) Date Survey Completed</b>  10/15/2018
<b>Name of Provider or Supplier</b>  South Georgia Family Medicine Associates	<b>Street Address, City, State</b>  1203 Brampton Ave, Statesboro, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on October 15, 2018. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiency was cited:
<b>D5429</b>	<p><b>MAINTENANCE AND FUNCTION CHECKS</b> CFR(s): 493.1254(a)(1)</p> <p>For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on observation by the surveyor during a tour of the laboratory and staff interview, the laboratory failed to perform maintenance and function checks on the LabCorp Horizon Mini E centrifuge used to prepare specimens for testing. Findings include: 1. Observation by the surveyor during a tour of the laboratory revealed a sticker on the centrifuge showing rotations per minute (RPM) check on the centrifuge was last performed on 5/10/2012.. 2. Interview with testing personnel (see CMS 209) on October 15,, 2018 at 12 pm in the laboratory confirmed there is no documentation of centrifuge maintenance and RPM check after 5/10/12 and maintenance is not performed every two years as recommended.</p>