Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D1091809	(X3) Date Survey Completed 10/17/2018
Name of Provider or Supplier Nuestros Ninos Our Kids Pediatrics	**Street Address, City, State** 777 Franklin Road, Marietta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on October 17, 2018. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D6015	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4) Ensure that the laboratory is enrolled in an HHS approved proficiency testing program for the testing performed. This STANDARD is not met as evidenced by: Based on proficiency testing (PT) document review and staff interviews, the lab director (LD) failed to ensure the laboratory is enrolled in a proficiency testing program. Findings include: 1. Review of American Proficiency Institute (API) documents revealed the lack of documents for 2018 testing events #1 and #2. 2. Interviews with staff #3 (CMS 209 form) and the LD on 10/17/18 at 11:14 am in the file/crash cart office confirmed the lab did not enroll in 2018 PT until 10/16/18.