

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D1107630	(X3) Date Survey Completed 06/20/2018
Name of Provider or Supplier Fasttrack Immediate Care	Street Address, City, State 1120-C Indain Drive, Eastman, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on June 20, 2018. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Performance Evaluation Review documents from American Proficiency Institute (API) Proficiency Testing (PT) provider, and staff interview the laboratory director was not reviewing and signing off on the PT Performance Evaluation Review. Findings: 1. Review of the PT performance evaluation review documents received from API, the review was not performed by the Laboratory Director or a qualified designee. 2018 1st Event Hematology was signed by an unqualified designee. 2017 3rd Event Hematology was signed by an unqualified designee 2. Interview with staff #2, (CMS 209 form on June 20, 2018 at approximately 1:45 pm in the facility office, confirmed that neither the LD or a qualified designee signed the performance Review and Corrective Action form for the above listed events.</p>
D5413	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(b)</p> <p>The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's</p>

instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.

This STANDARD is not met as evidenced by:
Based on the review of the Humidity Charts for the laboratory, and staff interview. The laboratory failed to document the Humidity in the lab from January 2018 to June 20, 2018. Findings: 1. Review of the Humidity charts for the laboratory the Humidity was not being monitored and recorded as required for the Abbott CelDyn Emerald Hematology analyzer. The range for the operation is listed as 20% to 80% 2. Interview with staff #2, (CMS209 FORMS), on June 20, 2018 at approximately 1:20pm in the office confirmed that the humidity was not being recorded, and had not been recorded since January 2018

D5791

ANALYTIC SYSTEMS QUALITY ASSESSMENT
CFR(s): 493.1289(a)(c)

(a) The laboratory must establish and follow written policies and procedures for an ongoing mechanism to monitor, assess, and when indicated, correct problems identified in the analytic systems specified in 493.1251 through 493.1283. (c) The laboratory must document all analytic systems assessment activities.

This STANDARD is not met as evidenced by:
Based on review of the Quality Assessment (QA) monthly monitor documents, and staff interview, the laboratory failed to perform QA checks and QA monitoring. Findings, 1.) Review of the QA procedure and QA monthly monitor documents, the laboratory had not completed a Monthly QA monthly monitor evaluation since January 2018 2. Interview with staff #2 (CMS) at approximately 12:50 pm in the facility office, confirmed that that were no Monthly QA monitors performed between January 2017 and June 2018.

D6003

LABORATORY DIRECTOR QUALIFICATIONS
CFR(s): 493.1405 AND 493.1406

The laboratory director must be qualified to manage and direct the laboratory personnel and the performance of moderate complexity tests and must be eligible to be an operator of a laboratory within the requirements of subpart R of this part. (a) The laboratory director must possess a current license as a laboratory director issued by the State in which the laboratory is located, if such licensing is required; and (b) The laboratory director must-- (b)(1)(i) Be a doctor of medicine or doctor of osteopathy licensed to practice medicine or osteopathy in the State in which the laboratory is located; and (b)(1)(ii) Be certified in anatomic or clinical pathology, or both, by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; or (b)(2)(i) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the Laboratory is located; and (b)(2)(ii) Have had laboratory training or experience consisting of: (b)(2)(ii)(A) At least one year directing or supervising non-waived laboratory testing; or (b)(2)(ii)(B) Beginning September 1, 1993, have at least 20 continuing medical education credit hours in laboratory practice commensurate

with the director responsibilities defined in 493.1407; or (b)(2)(ii)(C) Laboratory training equivalent to paragraph (b)(2)(ii)(B) of this section obtained during medical residency. (For example, physicians certified either in hematology or hematology and medical oncology by the American Board of Internal Medicine); or (b)(3) Hold an earned doctoral degree in a chemical, physical, biological, or clinical laboratory science from an accredited institution; and (b)(3)(i) Be certified by the American Board of Medical Microbiology, the American Board of Clinical Chemistry, the American Board of Bioanalysis, or the American Board of Medical Laboratory Immunology; or (b)(3)(ii) Have had at least one year experience directing or supervising non-waived laboratory testing; (b)(4)(i) Have earned a master's degree in a chemical, physical, biological or clinical laboratory science or medical technology from an accredited institution; (b)(4)(ii) Have at least one year of laboratory training or experience, or both in non-waived testing; and (b)(4)(iii) In addition, have at least one year of supervisory laboratory experience in non-waived testing; or (b)(5)(i) Have earned a bachelor's degree in a chemical, physical, or biological science or medical technology from an accredited institution; (b)(5)(ii) Have at least 2 years of laboratory training or experience, or both in non-waived testing; and (b)(5)(iii) In addition, have at least 2 years of supervisory laboratory experience in non-waived testing; (b)(6) Be serving as a laboratory director and must have previously qualified or could have qualified as a laboratory director under 493.1406; or (b)(7) On or before February 28, 1992, qualified under State law to direct a laboratory in the State in which the laboratory is located. Laboratory director qualifications on or before February 28, 1992 The laboratory director must be qualified to manage and direct the laboratory personnel and test performance. (a) The laboratory director must possess a current license as a laboratory director issued by the State, if such licensing exists; and (b) The laboratory director must: (b)(1) Be a physician certified in anatomical or clinical pathology (or both) by the American Board of Pathology or the American Osteopathic Board of Pathology or possess qualifications that are equivalent to those required for such certification; (b)(2) Be a physician who: (b)(2)(i) Is certified by the American Board of Pathology or the American Osteopathic Board of Pathology in at least one of the laboratory specialties; or (b)(2)(ii) Is certified by the American Board of Medical Microbiology, the American Board of Clinical Chemistry, the American Board of Bioanalysis, or other national accrediting board in one of the laboratory specialties; or (b)(2)(iii) Is certified by the American Society of Cytology to practice cytopathology or possesses qualifications that are equivalent to those required for such certification; or (b)(2)(iv) Subsequent to graduation, has had 4 or more years of full-time general laboratory training and experience of which at least 2 years were spent acquiring proficiency in one of the laboratory specialties; (b)(3) For the subspecialty of oral pathology only, be certified by the American Board of Oral Pathology, American Board of Pathology or the American Osteopathic Board of Pathology or possesses qualifications that are equivalent to those required for certification; (b)(4) Hold an earned doctoral degree from an accredited institution with a chemical, physical, or biological science as a major subject and (b)(4)(i) Is certified by the American Board of Medical Microbiology, the American Board of Clinical Chemistry, the American Board of Bioanalysis, or other national accrediting board acceptable to HHS in one of the laboratory specialties; or (b)(4)(ii) Subsequent to graduation, has had 4 or more years of full-time general laboratory training and experience of which at least 2 years were spent acquiring proficiency in one of the laboratory specialties; (b)(5) With respect to individuals first qualifying before July 1, 1971, have been responsible for the direction of a laboratory for 12 months between July 1, 1961, and January 1, 1968, and, in addition, either: (b)(5)(i) Was a physician and subsequent to graduation had at least 4 years of pertinent full-time laboratory experience; (b)(5)(ii) Held a master's degree from an accredited institution with a chemical, physical, or biological science

as a major subject and subsequent to graduation had at least 4 years of pertinent full-time laboratory experience; (b)(5)(iii) Held a bachelor's degree from an accredited institution with a chemical, physical, or biological science as a major subject and subsequent to graduation had at least 6 years of pertinent full-time laboratory experience; or (b)(5)(iv) Achieved a satisfactory grade through an examination conducted by or under the sponsorship of the U.S. Public Health Service on or before July 1, 1970; or (b)(6) Qualify under State law to direct the laboratory in the State in which the laboratory is located. Note: The January 1, 1968 date for meeting the 12 months' laboratory direction requirement in paragraph (b)(5) of this section may be extended 1 year for each year of full-time laboratory experience obtained before January 1, 1958 required by State law for a laboratory director license. An exception to the July 1, 1971 qualifying date in paragraph (b)(5) of this section was made provided that the individual requested qualification approval by October 21, 1975 and had been employed in a laboratory for at least 3 years of the 5 years preceding the date of submission of his qualifications.

This STANDARD is not met as evidenced by:
Based on review of the credentials of the listed Laboratory Director (LD), and staff interview, the laboratory does not meet this standard. Findings: 1. Review of the Laboratory Director's credentials and documentation, provided during the survey showed that the listed LD does not meet the qualifications to manage and direct the laboratory personnel and the performance of the moderate complexity tests. 2. Interview with staff #2 (CMS form 209), on June 20, 2018 at approximately 2 pm in the office, confirmed that the only documents that she had was what was given to me during the survey. 3. I requested more documentation from staff #2(CMS form 209) consisting of any laboratory training, and no other documentation was provided as of July 3, 2018

D6004

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1407(a)(b)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (a) The laboratory director, if qualified, may perform the duties of the technical consultant, clinical consultant, and testing personnel, or delegate these responsibilities to personnel meeting the qualifications of 493.1409, 493.1415, and 493.1421, respectively. (b) If the laboratory director reappoints performance of his or her responsibilities, he or she remains responsible for ensuring that all duties are properly performed.

This STANDARD is not met as evidenced by:
Based on review of the Proficiency Testing (TP) documents, Quality Control (QC) documents, Maintenance Logs, Temperature/Humidity Charts, Quality Assessment (QA) documents the Laboratory Director failed to oversee the overall operation and administration of the laboratory. Findings: 1. Review of the PT documents showed that the LD had not signed Evaluation Reports and approved/recommended any correction action that was needed. QC documents have not been reviewed and corrective action recommended from January 2017 to June 2018. Maintenance Logs have not been documented as reviewed from January 2017 to June 2018. QA documents were not being performed and monitored by the LD from January 2017 to

June 2018. There is no document of Monthly Visits as well as educational presentations by the LD from January 2017 to June 2018. 2. Interview with staff #2, (CMS form 209) on June 20, 2018 at approximately 2pm, in the facility office, confirmed that there has not been LD oversight since January 2017 to June 2018.