

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2004759	(X3) Date Survey Completed 09/13/2022
Name of Provider or Supplier D Conrad Harper Md, Llc	Street Address, City, State 101 Nw Bowens Mill Rd, Suite D, Douglas, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) Recertification survey was completed on September 13, 2022. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3) -- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.</p> <p>This STANDARD is not met as evidenced by: Based on review of the calibration documents for the Beckman/Coulter DH50, and</p>

staff interview, the laboratory failed to perform calibration at every 6 months as required. Findings: 1. Review of the calibration documents for the Beckman Coulter DH50, the laboratory had calibration documents for November 2020 and August 2022. There were no documented calibrations between those dates. 2. Interview with staff #2 (CMS 209 Personnel Form) on 09/13/2022 at approximately 2pm in the breakroom, confirmed the aforementioned statement. An email received on 9/16/2022, at 12:13pm, confirmed the lack of documentation to support the required calibrations.

D6005

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1407(c)

The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (c) The laboratory director must be accessible to the laboratory to provide onsite, telephone or electronic consultation as needed.

This STANDARD is not met as evidenced by:
Based on review of the calibration documents for the Beckman/Coulter DH50, and staff interview, the Laboratory Director failed to provide oversight for the laboratory. The laboratory failed to perform Calibration on the DH50 every 6 months as required. Findings: 1. Review of the calibration documents for the Beckman Coulter DH50, the laboratory had calibration documents for the following dates: November 2020, and August 2022. There were no documented calibrations between those dates. 2. Interview with staff #2 (CMS 209 Personnel Form) on 09/13/2022 at approximately 2pm in the breakroom, confirmed the aforementioned statement. An email was received on 9/16/ 2022 at 12:13pm confirmed the lack of oversight by the Laboratory Director to ensure completion of all required calibrations.