

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2017911	(X3) Date Survey Completed 11/11/2025
Name of Provider or Supplier Back To Life Medical Group, Llc	Street Address, City, State 557 Riverstone Parkway, Suite 140, Canton, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) Recertification Survey was completed on November 11, 2025. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5415	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(c)</p> <p>(c) Reagents, solutions, culture media, control materials, calibration materials, and other supplies, as appropriate, must be labeled to indicate the following: (c)(1) Identity and when significant, titer, strength or concentration. (c)(2) Storage requirements. (c)(3) Preparation and expiration dates. (c)(4) Other pertinent information required for proper use.</p> <p>This STANDARD is not met as evidenced by: A tour of the laboratory revealed that testing personnel failed to label, in use reagents, with an opened date and/ or expiration date. THE FINDINGS INCLUDE: 1. A tour of the laboratory confirmed that controls and reagents, in use for testing, were not dated with an opened date and/ or an expiration date as required. 2. An exit interview, with the Testing Personnel, Office Manager and the Laboratory Director via phone conference, on November 11, 2025, at 2:30pm, confirmed that the laboratory had multiple reagents in use without labeled opened and/or expiration date.</p>
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p>

This STANDARD is not met as evidenced by:

A tour of the laboratory revealed that the laboratory had multiple expired reagents in use for clinical laboratory testing. THE FINDINGS INCLUDE: 1. A tour of the laboratory revealed the following expired reagents: a. 3 boxes of ThermoFisher Scientific MGC Clinical DAU Control Set. 1 box expired on 06/30/2024 and 2 boxes expired on 09/30/2025; b. BD EDTA Vacutainer Tubes that expired 08/31/2020; c. BD Urinalysis Urine Vacutainer Tubes that expired on 08/31/2020; d. Grainer Bio-One GmbH Vacuette Tubes that expired on 09/13/2025; and e. 1 vial of ThermoFischer Scientific Washing Solution 4.5% that expired on 06/20/2024. 2. An exit interview with Testing Personal and Office Manager, as well as the Laboratory Director via phone conference, on November 11, 2025, at 2:30pm confirmed that the laboratory had multiple expired reagents and supplies, on site, in use for clinical laboratory testing.