

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2025061	(X3) Date Survey Completed 01/12/2021
Name of Provider or Supplier Georgia Dermatology & Skin Cancer Center	Street Address, City, State 2283 Wrightsboro Road, Augusta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on January 12, 2021. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5449	<p>CONTROL PROCEDURES CFR(s): 493.1256(d)(3)(ii)(g)</p> <p>Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- At least once a day patient specimens are assayed or examined perform the following for-- Each qualitative procedure, include a negative and positive control material; (g) The laboratory must document all control procedures performed.</p> <p>This STANDARD is not met as evidenced by: Based on patient document review and staff interview, the laboratory failed to perform qualitative quality control (QC) at least once a day patient specimens were tested as required. Findings include: 1. Patient document review revealed QC was not performed for Potassium Hydroxide (KOH) and Wet Preparation (Parasitology) for the following time periods: 2018 (July - December), 2019, 2020, and 2021 thus far. 2. An interview with the clinic administrator in the upstairs breakroom on 1/12/2021 at approximately 2:30 p.m. confirmed the lack of QC performed for KOH and Wet Preparation for the aforementioned time periods.</p>