

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2047071	(X3) Date Survey Completed 08/05/2026
Name of Provider or Supplier Cole Dermatology, Llc	Street Address, City, State 2410 North Oak Street, Valdosta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on August 05, 2026. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D6128	TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9) (b)(9) Thereafter, evaluations must be performed at least annually unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individuals performance must be reevaluated to include the use of the new test methodology or instrumentation. This STANDARD is not met as evidenced by: Based on review of testing personnel documents and an interview with the lab director, the technical supervisor failed to perform annual competency on all testing personnel. Findings: 1. Review of personnel competency records revealed the techincal supervisor failed to perform the annual competency on 1 of 5 testing personnel. No competency records were presented for testing personnel #2 (CMS 209) for the year 2025 and 2026 to date. 2. Interview with the lab director(CMS 209) confirmed the aforementioned findings on August 05, 2026 at 11:23 am in the breakroom.