

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2053177	(X3) Date Survey Completed 07/09/2019
Name of Provider or Supplier K & S Clinical Diagnostics	Street Address, City, State 1995 North Park Pl, Se, Suite 205, Atlanta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on July 8, 2019. The laboratory was found not in compliance with all applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: A review of the procedure manual and interview with the Technical Supervisor (TS), determined that the laboratory failed to have a comprehensive competency assessment for their testing personnel for the specialty of Toxicology panel from November 2017 through July 2019. The findings include: 1. A review of testing personnel records revealed that a competency assessment was performed on testing personnel, but the assessment was not specific to the laboratory's toxicology panel testing from November 2017 through July 2019. 2. The laboratory must monitor each individual's competency and identify remedial training for each test procedure performed in the laboratory. 3. The TS confirmed on July 8, 2019 at 12:45 PM, in the office, that the laboratory failed to have a competency assessment that was specific to their toxicology panel testing.</p>