

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2061013	(X3) Date Survey Completed 11/24/2025
Name of Provider or Supplier Wellstar Urgent Care Crabapple	Street Address, City, State 12460 Crabapple Road, Suite 901, Alpharetta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review was completed on November 24, 2025. At the time of the review, the laboratory was not in compliance with the Clinical Laboratory Improvement Amendments of 1988, 42 CFR 493.1 through 42 CFR 493.1780. The following condition deficiencies were cited: D2016 - 42 CFR 493.803 Condition: Successful participation [proficiency testing] D6000 - 42 CFR 493.1403 Condition: Moderate Complex Laboratory Director
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the CASPER 155 report and review of the American Proficiency Institute (API) reports, the laboratory failed to maintain satisfactory proficiency</p>

	<p>testing (PT) participation for hemaglobin (HGB) in 2025 events 1 & 3, resulting in an initial unsuccessful participation for HGB. Refer to D 2130</p>
D2130	<p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>(f) Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Centers for Medicare and Medicaid (CMS) CASPER 155 report and review of American Proficiency Institute (API) reports, the laboratory failed to maintain satisfactory participation in two of three testing events (2025 events 1 & 3), resulting in an initial unsuccessful participation for HGB. Findings: 1. A review of Casper Report 155 revealed the laboratory failed HGB on the following: 2025 Event 1 HGB Score 60% 2025 Event 3 HGB Score 60% 2. A review of the laboratory's API Reports confirmed the laboratory failed HGB with the aforementioned scores.</p>
D6000	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on review of the CMS CASPER 155 report and review of American Proficiency Institute (API) reports, the laboratory director failed to provide overall management and direction for proficiency testing performance. The laboratory director failed to ensure proficiency testing samples were tested as required. Refer to D6016</p>
D6016	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>(e)(4)(i) The proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on review of the CMS CASPER Report 155 and the American Proficiency Institute (API) 2025 events 1 & 3 PT evaluation reports, the laboratory director failed to ensure successful proficiency testing performance in HGB in two of three testing events (2025 events 1 and 3), resulting in the initial unsuccessful participation for HGB. Findings: 1. A review of Casper Report 155 revealed the laboratory failed HGB on the following: 2025 Event 1 HGB Score 60% 2025 Event 3 HGB Score 60% 2. A review of the laboratory's API Reports confirmed the laboratory failed HGB with the aforementioned scores.</p>