

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2084143	(X3) Date Survey Completed 03/18/2021
Name of Provider or Supplier Piedmont Cancer Institute	Street Address, City, State 1240 Eagles Landing Parkway, Stockbridge, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on March 18, 2021. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D6054	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least annually, after the first year.</p> <p>This STANDARD is not met as evidenced by: Based on testing personnel (TP) competency document review and laboratory supervisor's interview, the laboratory director (LD) failed to ensure TP annual competencies were performed by qualified staff as required. Findings include: 1. TP document review revealed annual competencies for TPs# 1 - 3 and #5 (CMS 209) were not performed by the Technical Consultant(TC) who is also the laboratory director (LD) in 2019 and 2020. 2. An interview with the laboratory supervisor, TP #2 (CMS209) on 03/18/2021 in a break room room at approximately 12:25 p.m. confirmed the aforementioned annual competencies were not performed in 2019 and 2020 by the (TC) who is also the laboratory director.</p>