

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2091367	(X3) Date Survey Completed 08/09/2022
Name of Provider or Supplier Select Interventional Pain, Pc	Street Address, City, State 794 McDonough Road, Suite 108, Jackson, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on August 09, 2022. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on proficiency testing (PT) document review and staff interview, the laboratory failed to verify at least twice annually the accuracy of pH testing performed. Findings include: 1. Proficiency testing (PT) document review revealed there was no twice annual verification of the accuracy of pH performed on the Global 240 analyzer. 2. An interview with staff #2 (CMS 209) in room #8 on 08/09/2022 at approximately 12:05 p.m. confirmed the lack of twice annual verification for the pH analyte.</p>
D5429	<p>MAINTENANCE AND FUNCTION CHECKS CFR(s): 493.1254(a)(1)</p> <p>For unmodified manufacturer's equipment, instruments, or test systems, the laboratory must perform and document maintenance as defined by the manufacturer and with at least the frequency specified by the manufacturer.</p> <p>This STANDARD is not met as evidenced by: Based on observation and interview with testing personnel (TP), the lab failed to calibrate the Champion F33D centrifuge (SN F33D1004782) per the procedure</p>

manual. Findings include: 1. Observation during the lab tour on 8/9/22 at 1035 AM revealed the Champion F33D centrifuge (SN F33D1004782) was last calibrated by the manufacturer in 2019 upon purchase. No documentation of calibration for 2020, 2021, or 2022 to date. 2. Interview with TP #2 (CMS 209 form) on 08/09/2022 in the lab at approximately 10:35 AM, confirmed the centrifuge had not been calibrated since 2019.