

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2118976	(X3) Date Survey Completed 05/26/2022
Name of Provider or Supplier Urgent Care Of Oconee - Watkinsville	Street Address, City, State 2061 Experiment Station Road, Suite 505, Watkinsville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on May 26, 2022. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5311	<p>SPECIMEN SUBMISSION, HANDLING, AND REFERRAL CFR(s): 493.1242(a)</p> <p>The laboratory must establish and follow written policies and procedures for each of the following, if applicable: (1) Patient preparation. (2) Specimen collection. (3) Specimen labeling, including patient name or unique patient identifier and, when appropriate, specimen source. (4) Specimen storage and preservation. (5) Conditions for specimen transportation. (6) Specimen processing. (7) Specimen acceptability and rejection. (8) Specimen referral.</p> <p>This STANDARD is not met as evidenced by: Based on review of the general laboratory standard operating procedure manual (SOP) and staff interview, the laboratory failed to establish written instructions for sending specimens to an outside reference laboratory for testing The findings include: 1. The SOP did not include a written policy and procedure (to include collection, preservation, storage, transport, testing schedule times, or how to obtain additional assistance) for staff to follow when sending specimens to reference laboratory (Quest, and LabCorp). 2. During an interview on May 26, 2022 at: 11:45 AM with Testing Personnel #1(CMS 209), in an exam room in the back of the facility, confirmed that the laboratory did not have a written policy and procedure for staff to follow when sending specimens to reference laboratories.</p>