

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2118976	(X3) Date Survey Completed 04/17/2024
Name of Provider or Supplier Urgent Care Of Oconee - Watkinsville	Street Address, City, State 2061 Experiment Station Road, Suite 505, Watkinsville, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A proficiency testing desk review was completed on April 17, 2024. At the time of the review, the laboratory was not in compliance with the Clinical Laboratory Improvement Amendments of 1988, 42 CFR 493.1 through 42 CFR 493.1780. The following condition deficiencies were cited: D2016 - 42 CFR 493.803 Condition: Successful participation [proficiency testing D6000 - 42 CFR 493.1403 Condition: Moderate Complex Laboratory Director
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on review of the Centers for Medicare and Medicaid (CMS) CASPER 155</p>

	<p>report, the laboratory failed to successfully participate in proficiency testing (PT) for hematocrit (HCT) & Hemaglobin (HGB) resulting in the non-initial unsuccessful participation for HCT & HGB. Refer to D 2130</p>
D2130	<p>HEMATOLOGY CFR(s): 493.851(f)</p> <p>Failure to achieve satisfactory performance for the same analyte in two consecutive events or two out of three consecutive testing events is unsuccessful performance.</p> <p>This STANDARD is not met as evidenced by: Based on review of the CMS CASPER 155 report, the laboratory failed to demonstrate satisfactory performance in 3 out of 3 consecutive testing events for HGB & HCT, resulting in the non- initial unsuccessful participation for HGB & HCT. Findings: 1. A review of Casper Report 155 revealed the laboratory failed HGB on the following: 2022 Event 2 Score 0% 2022 Event 3 Score 0% 2023 Event 1 Score 0% 2. A review of Casper Report 155 revealed the laboratory failed HCT on the following: 2022 Event 2 Score 0% 2022 Event 3 Score 0% 2023 Event 1 Score 0%</p>
D6000	<p>MODERATE COMPLEXITY LABORATORY DIRECTOR CFR(s): 493.1403</p> <p>The laboratory must have a director who meets the qualification requirements of 493.1405 of this subpart and provides overall management and direction in accordance with 493.1407 of this subpart.</p> <p>This CONDITION is not met as evidenced by: Based on review of the CMS CASPER 155 report, the laboratory director failed to provide overall management and direction for successful participation in PT. The laboratory director failed to ensure PT samples were tested as required. Refer to D6016</p>
D6016	<p>LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(4)(i)</p> <p>The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(4)(i) Ensure that the proficiency testing samples are tested as required under Subpart H of this part;</p> <p>This STANDARD is not met as evidenced by: Based on review of the CMS CASPER Report 155, the laboratory director failed to ensure successful PT participation in 3 out of 3 consecutive testing events. Refer to D 2130</p>