

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  11D2133715	<b>(X3) Date Survey Completed</b>  11/26/2025
<b>Name of Provider or Supplier</b>  3 16 Family Medicine Kids Coner	<b>Street Address, City, State</b>  107 Wedgewood Drive, Carrollton, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	A Clinical Laboratory Improvement Amendments (CLIA) Recertification survey was completed on November 26, 2025 . The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
<b>D5221</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(d)</p> <p>All proficiency testing evaluation and verification activities must be documented.</p> <p>This STANDARD is not met as evidenced by: Based on review of American Proficiency Institute (API) reports and an interview with the technical consultant (TC), the laboratory failed to document corrective actions for unsuccessful Proficiency Testing (PT) scores ( 1st and 3rd events of 2024). Findings: 1. A review of API result reports revealed the laboratory failed Sodium (NA) on the following: 2024 Event 1 NA Score 60% 2024 Event 3 NA Score 60% 2. An interview with the TC (CMS 209) on 11/26/25 at 11:30 am in the laboratory office area confirmed the laboratory failed NA with the aforementioned scores without documented corrective actions.</p>