Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2134965	(X3) Date Survey Completed 01/08/2025
Name of Provider or Supplier Wellstar Vining's Health Park Urgent Care	**Street Address, City, State** 4441 Atlanta Road Se, Suite 107, Smyrna, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on January 1/8/25. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D6029	LABORATORY DIRECTOR RESPONSIBILITIES CFR(s): 493.1407(e)(11) The laboratory director is responsible for the overall operation and administration of the laboratory, including the employment of personnel who are competent to perform test procedures, and record and report test results promptly, accurate, and proficiently and for assuring compliance with the applicable regulations. (e) The laboratory director must-- (e)(11) Ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results. This STANDARD is not met as evidenced by: Based on review of the training & competency personnel documents and interview with a technical consultant (TC), the lab director (LD) failed to ensure all testing personnel (TP) were trained at each laboratory that they perform laboratory technical procedures. Findings: 1. Review of the training documents provided revealed the float and the PRN (as needed) TP (CMS form 209 page 2 of 3), 9 of 9 TP listed were trained under the location of WMG not the Wellstar Urgent Care Vinings Health Park (VHP) location. 2. CLIA regulations requires the completion of training and competency of all testining personnel per CLIA Certificate. 3. Interview with TC #3 (CMS 209 form page 3 of 3) on 1/8/25 at 1:30 pm, confirmed the aforementioned TP were trained under the WMG location to work at any Wellstar Medical Group laboratory.