

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2141045	(X3) Date Survey Completed 03/17/2021
Name of Provider or Supplier Adma Bio Centers Georgia, Inc	Street Address, City, State 166 Earnest W Barrett Parkway Nw, Marietta, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on March 17, 2021. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on testing personnel (TP) document review and staff interview, the laboratory failed to establish and follow written policies and procedures to assess TP competency as required. Findings include: 1. TP competency document review revealed the laboratory failed to include the six required CLIA procedures in performance of TP competencies. 2. An interview with the Center Director on 3/17/2021 in the conference room at approximately 12:30 p.m. confirmed TP competency performance did not include the aforementioned CLIA procedures.</p>