

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 11D2224642	(X3) Date Survey Completed 01/06/2023
Name of Provider or Supplier Adma Biocenters Georgia, Inc	Street Address, City, State 1820 Highway 20, Se, Suite 112, Conyers, GA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A Clinical Laboratory Improvement Amendments (CLIA) recertification survey was completed on January 6, 2023. The laboratory was not in compliance with applicable CLIA requirements found at 42 CFR 493.1 through 42 CFR 493.1780. The following deficiencies were cited:
D3011	<p>FACILITIES CFR(s): 493.1101(d)</p> <p>Safety procedures must be established, accessible, and observed to ensure protection from physical, chemical, biochemical, and electrical hazards, and biohazardous materials.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Standard Operating Procedure (SOP) and staff interview, the laboratory failed to establish a safety procedure for the Eyewash. The Findings include: 1. SOP document review revealed that the laboratory failed to establish a safety procedure for the sink eyewash in the laboratory. 2. During an interview with the Center Director, Center Quality Manager, and the Center Quality Assurance Director (CMS-209) on January 6, 2023, at approximately 1:45 PM, in the conference room upstairs, in the facility and also in the processing area of the donor center, confirmed that the laboratory failed to establish a safety procedure for the Eyewash.</p>