

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  13D0520334	<b>(X3) Date Survey Completed</b>  04/23/2024
<b>Name of Provider or Supplier</b>  Family Practice Group, Pa	<b>Street Address, City, State</b>  1951 Bench Rd #B, Pocatello, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5211</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of proficiency testing (PT) documentation from the American Proficiency Institute (API) and an interview with the technical consultant (TC) on 4/23/2024, the laboratory director failed to review PT results in 2022, 2023 and 2024. The findings include: 1. A review of PT documents for 2022 from API identified that the laboratory director failed to review and evaluate results for urine microscopic examination for events two and three. 2. A review of PT documents for 2023 from API identified that the laboratory director failed to review and evaluate results for urine microscopic examination for events one, two and three. 3. A review of PT documents for 2024 from API identified that the laboratory director failed to review and evaluate results for urine microscopic examination for event one. 4. An interview with the TC on 4/23/2024 at 8:51 am confirmed that the laboratory director failed review the above PT results. 5. The laboratory reports performing 1842 urine microscopic examinations annually.</p>