

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0520814	(X3) Date Survey Completed 08/13/2021
Name of Provider or Supplier Pediatric Center	Street Address, City, State 3430 Washington Pkwy, Idaho Falls, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a record review of proficiency testing (PT) from American Proficiency Institute (API) and an interview with the Technical Supervisor (TS) on 8/13/2021, the laboratory failed to review PT and evaluate results that were less than 100% but greater than or equal to 80% for 2020 and 2021. The findings include: 1. A review of PT records from API for Microbiology 2020 event two identified that the laboratory failed to evaluate the results for gram stains with a score of 80% due to an unacceptable result for sample GS-09. 2. A review of PT records from API for Hematology/Coagulation 2020 event three identified that the laboratory failed to evaluate the results for Red Cell Count and MCH, performed on the Horiba ABX Micros 60, with a score of 80% due to unacceptable results for sample HEM-13. 3. A review of PT records from API for Hematology/Coagulation 2021 event one identified that the laboratory failed to evaluate the results for Hematocrit, MCHC, MCV and RDW, performed on the Horiba ABX Micros 60, with a score of 80% due to unacceptable results for sample HEM-01. 4. An interview with the TS on 8/13/2021 at 10:00 am confirmed that the laboratory failed to review and evaluate PT results from API that were less than 100% but greater than or equal to 80% for 2020 and 2021. 5. The laboratory reports performing 13,564 moderate and high complexity tests annually.</p>