Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0520888	(X3) Date Survey Completed 04/27/2023
Name of Provider or Supplier Urology Associates Of Idaho Falls	**Street Address, City, State** 2375 Coronado St, Idaho Falls, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235 As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency. This STANDARD is not met as evidenced by: Based on a review of the Centers for Medicare and Medicaid Services (CMS) 209 personnel form, competency assessments and an interview with the technical consultant (TC) on 4/27/2023, the laboratory failed to have documentation of annual competency for two (2) of 10 testing personnel in 2022. The findings include: 1. A review of competency assessments identified that the laboratory failed to have annual competency assessment for one testing personnel (TP1) performing testing on the NanoEntek FREND and and for one testing personnel (TP2) performing urine microscopic examinations listed on the CMS 209 for 2022. 2. An interview with the TC on 4/27/2023 at 1:09 pm confirmed that there was no annual competency assessment for TP1 and TP2 for 2022. 3. The laboratory reports performing 2783 tests annually.