

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0520943	(X3) Date Survey Completed 03/25/2026
Name of Provider or Supplier Obgyn Associates Of Idaho Falls	Street Address, City, State 2065 E 17th St #D, Idaho Falls, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the submitted Centers for Medicare and Medicaid Services (CMS) 209 personnel form, laboratory policies, competency assessment records and an interview with the operations manager on 3/25/2026, the laboratory failed to establish and follow written policies and procedures to assess testing personnel training and competency in 2024 and 2025. The findings include: 1. The CMS 209 identified four (4) testing personnel performing moderate complexity testing. 2. A review laboratory policies identified the laboratory failed to have a policy for performing training and competency assessments. 3. A review of training and competency assessment records identified four (4) of four (4) testing personnel that failed to have annual competency assessment in 2024 and 2025. 4. An interview with the operations manager on 3/25/2026 at 9:25 am confirmed the above findings. 5. The laboratory reports performing 1275 tests annually.</p>