

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0521272	(X3) Date Survey Completed 05/17/2021
Name of Provider or Supplier Valor Health	Street Address, City, State 1202 E Locust St, Emmett, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on a Proficiency Testing (PT) desk review of the Centers for Medicare and Medicaid (CMS) PT data report (Report 155D) and graded results from the College of American Pathologists (CAP) the laboratory failed to successfully participate in PT for the specialty of Routine Chemistry for two (2) of three (3) consecutive testing events. The findings include: 1. A PT desk review revealed the laboratory had unsatisfactory scores for the 3rd event of 2020 and the 1st event of 2021 for the specialty of Routine Chemistry.
D2096	ROUTINE CHEMISTRY

CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a Proficiency Testing (PT) desk review of the Centers for Medicare and Medicaid (CMS) PT data report (Report 155D) and graded results from the College of American Pathologists (CAP), the laboratory failed to achieve an overall testing event score of satisfactory performance for two (2) consecutive testing events for the specialty of Routine Chemistry for the analyte BUN. The findings include: 1. A PT desk review of report 155D and graded results from CAP revealed the laboratory failed to achieve satisfactory results for the specialty of Routine Chemistry for the analyte BUN. Analyte Year Event Score BUN 2020 3 20 BUN 2021 1 40