

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0521865	(X3) Date Survey Completed 06/09/2022
Name of Provider or Supplier Dermatology Clinic Of Idaho	Street Address, City, State 7733 W Emerald St, Boise, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on a review of laboratory procedures and an interview with the laboratory director on 6/9/2022, the laboratory failed to have a procedure for potassium hydroxide (KOH) testing available for testing personnel to follow. The finding include: 1. A review of laboratory procedures identified that the laboratory failed to have a procedure for KOH testing available to testing personnel since beginning testing in February 2022. 2. An interview with the laboratory director on 6/9/2022 at 10:06 am confirmed that the laboratory failed to have a procedure for KOH testing. 3. The laboratory reports performing 25 KOH tests annually.</p>