

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0646318	(X3) Date Survey Completed 01/28/2022
Name of Provider or Supplier Glenns Ferry Health Center Inc	Street Address, City, State 120 Desert Sage Way, Mountain Home, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a record review of proficiency testing (PT) from the American Proficiency Institute (API) and an interview with the laboratory lead on 1/28/2022, the laboratory failed to review and evaluate PT results that were less than 100% but were equal to or greater than 80%. The findings include: 1. A review of PT records from API for 2020 event two identified that the laboratory failed to evaluate results for the analyte white blood count (WBC), which had a score of 80%. 2. An interview with the laboratory lead on 1/28/2022 at 10:42 am confirmed that the laboratory failed to evaluate the PT WBC results for 2020 event two. 3. The laboratory reports performing 245 hematology tests annually.</p>