

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  13D0669093	<b>(X3) Date Survey Completed</b>  04/25/2018
<b>Name of Provider or Supplier</b>  Marimn Health	<b>Street Address, City, State</b>  427 N 12th St, Plummer, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on an observation of reagents in the lab and an interview with the laboratory manager, the laboratory had expired potassium hydroxide (KOH) solution in use for the detection of fungal elements in patient samples since October 2017. Findings: 1. An observation of the KOH solution by the microscope in the lab, revealed two HealthLink KOH bottles with lot number 1603217 and expiration date of 2017/10/31. 2. An interview on April 25, 2018 at 11:00 AM, with the laboratory manager, confirmed the KOH used in patient testing was expired.</p>