

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0688492	(X3) Date Survey Completed 06/07/2023
Name of Provider or Supplier Snake River Dermatology	Street Address, City, State 2101 N Whitley Dr, Fruitland, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of bi-annual verification documentation and an interview with the laboratory technicians on 6/07/2023, the laboratory failed to document, at least twice annually, the verification of accuracy for potassium hydroxide (KOH) examinations and Mohs surgical procedures in 2022 and 2023. The findings include: 1. A review of documentation for bi-annual verification of KOH examinations identified that the laboratory failed to document verification of accuracy for KOH for 2022 and 2023. 2. A review of documentation for bi-annual verification of Mohs surgical procedures identified that the laboratory failed to document verification of accuracy for Mohs surgical procedures for 2022. 3. An interview with the laboratory technicians on 6/07/2023 at 9:39 am confirmed the above findings. 3. The laboratory reports performing 1500 KOH examinations and 600 Mohs surgical procedures annually.</p>