

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 13D0688492 | (X3) Date Survey Completed 06/18/2025 |
| Name of Provider or Supplier Snake River Dermatology | Street Address, City, State 2101 N Whitley Dr, Fruitland, ID | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|---|
| D5217 | <p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a lack of documentation and an interview with the laboratory lead on 6/18 /2025, the laboratory failed to document, at least twice annually, the verification of accuracy for potassium hydroxide (KOH) examinations in 2023 and 2024. The findings include: 1. A lack of documentation for bi-annual verification once in 2023 and twice in 2024 identified that the laboratory failed to document verification of accuracy for KOH examinations at least twice annually in 2023 and 2024. 2. An interview with the laboratory lead on 6/18/2025 at 10:00 am confirmed that the laboratory failed to perform bi-annual verification of KOH examinations in 2023 and 2024. 3. The laboratory reports performing 150 KOH examinations annually</p> |
| D5417 | <p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on a direct observation and an interview with the laboratory lead on 6/18/2025, the laboratory failed to discontinue the use of expired urine Human Chorionic Gonadotropin (hCG) tests in 2023. The findings include: 1. A direct observation of the laboratory's hCG testing supplies on 6/18/2025 identified that the laboratory failed</p> |

to discontinue the use of 6 McKesson Consult urine hCG tests, expiration 7/31/2023 prior to the expiration. 2. An interview with the laboratory lead on 6/18/2025 at 10:50 am confirmed the above finding. 3. The laboratory reports performing 800 urine hCG tests annually