

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 13D0703662	(X3) Date Survey Completed 10/15/2024
Name of Provider or Supplier Cascade Medical Center	Street Address, City, State 402 Lake Cascade Pkwy, Cascade, ID	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on a proficiency testing (PT) desk review of the Centers for Medicare and Medicaid (CMS) PT data report (Report 155D), graded results from the American Proficiency Institute (API) and a telephone interview with the laboratory manager on 10/10/2024 at 9:35 AM, the laboratory failed to successfully participate in proficiency testing for the specialty routine chemistry for the analyte cholesterol, total 2 (two) out of 3 (three) events for 2024. Refer to: D2096
D2096	ROUTINE CHEMISTRY CFR(s): 493.841(f)

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on a review of the Centers for Medicare and Medicaid (CMS) Proficiency Testing (PT) data report (Report 155D), graded PT results from the American Proficiency Institute (API), and an interview with the laboratory manager on 10/10/24, the laboratory failed to achieve satisfactory performance for two (2) out of three (3) PT events for the analyte cholesterol, total. The findings include: 1. A review of Report 155D and graded PT results from API identified that the laboratory failed to achieve satisfactory performance for events two (2) and (3) three in 2024 for the specialty of routine chemistry for the analyte cholesterol, total. Provider Analyte Year Event Score API cholesterol, total 2024 2 40% API cholesterol, total 2024 3 60% 2. An interview with the laboratory manager on 10/10/2024 at 9:35 AM confirmed the above findings.